NOTES:

- 1. SUBSTRATE UV GRADE FUSED SILICA
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 ARCSEC
- 3. COATING (APPLY ACROSS COATING APERTURE)


PERFORMANCE SPECIFICATIONS @ 0° AOI T(AVG): >85% FROM 1290 - 2000nm T(AVG): < 1% FROM 960 - 1200nm T(ABS): = 50% @ 1250nm

S1: MULTILAYER DIELECTRIC COATING S2: SINGLE LAYER MgF2



- 5. TRANSMITTED WAVEFRONT DISTORTION RMS@ 632.8nm ≤ 1/4 WAVE
- 6. ROHS COMPLIANT





DO NOT

LONGPASS	FILTER
----------	--------

Т

	S1	\$2				<sup>©</sup> Edmund Optic	<b>C</b> ®
SHAPE	PLANO	PLANO					5
SURFACE QUALITY	40-20	40-20				OD2, Ø12.5mm, 1250nm LONGPASS F	
CLEAR APERTURE	80%	80%	THIRD ANGLE . PROJECTION	$\bigcirc \bigcirc$	TITLE		
COATING APERTURE	Ø11.00	Ø11.00					CULLET
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN	mm	DWG NO	84665	SHEET 1 OF 1