## NOTES:

1. SUBSTRATE MATERIAL:

N-BK7

2. COATING (APPLY ACROSS CLEAR APERTURE)

S1 & S2: VIS-NIR

R(abs) ≤0.25% @ 880nm


R(avg) ≤1.25% @ 400 - 870nm


R(avg) ≤1.25% @ 890 - 1000nm

3. FINE GROUND SURFACE

4. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICAIONS APPLY ACROSS CLEAR APERTURE

5. Rohs Compliant





## FOR INFORMATION ONLY: PARTS TO THIS DRAWING

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

REV. A	S1	S2				
SHAPE	CONVEX	PLANO	EFL	75		Edmund Ontice
RADIUS	38.76	∞	BFL	73.68		Edmund Optics®
SURFACE QUALITY	60-40	60-40		1		10mm DIA X 75mm FL VIS-NIR COATED,
CLEAR APERTURE	Ø9.00	Ø9.00	THIRD ANGLE PROJECTION		TITLE	CYLINDER LENS
COATING APERTURE	Ø9.00	Ø9.00		I		
BEVEL	PROTECTED AS NEEDED	PROTECTED AS NEEDED	ALL DIMS IN	mm	DWG NO	47965 SHEET 1 OF