NOTES:

1. SUBSTRATE MATERIAL:

N-BK7

2. COATING (APPLY ACROSS CLEAR APERTURE)

S1 & S2: VIS-NIR

R(abs) ≤0.25% @ 880nm


R(avg) ≤1.25% @ 400 - 870nm


R(avg) ≤1.25% @ 890 - 1000nm

3. FINE GROUND SURFACE

4. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICAIONS APPLY ACROSS CLEAR APERTURE

5. Rohs Compliant





FOR INFORMATION ONLY:
DO NOT MANUFACTURE PARTS TO THIS DRAWING

> SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

REV. A	\$1	S2				
SHAPE	CONVEX	PLANO	EFL	30		Edmund Ontion
RADIUS	15.5	∞	BFL	28.02		Edmund Optics®
SURFACE QUALITY	60-40	60-40				5mm H X 10mm L X 30mm FL VIS-NIR COATED,
CLEAR APERTURE	4.00 X 9.00	4.00 X 9.00	THIRD ANGLE PROJECTION		TITLE	CYLINDER LENS
COATING APERTURE	4.00 X 9.00	4.00 X 9.00		1		
BEVEL	PROTECTED AS NEEDED	PROTECTED AS NEEDED	ALL DIMS IN	mm	DWG NO	47959 SHEET 1 OF 1