- 1. SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)
  - \$1: HARD DIELECTRIC SPUTTERED T(avg): ≥91% FROM 408 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 390nm @ 0° AOI T(abs): =50% FOR 400±4nm @ 0° AOI

S2:SINGLE LAYER MgF2

### 4. FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT







6



Τ

			_		Edmund Ontice®	
REV A	S1	\$2				
SHAPE	PLANO	PLANO			Ø25mm 400nm HIGH PERFORMANCE	
SURFACE QUALITY	40-20	40-20		TITLE	IONGPASS FILTER	
CLEAR APERTURE	>80%	>80%				
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN mm	DWG NO	62981 SHEET	1

- 1. SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)
  - \$1: HARD DIELECTRIC SPUTTERED T(avg): ≥91% FROM 458 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 440nm @ 0° AOI T(abs): =50% FOR 450±4.5nm @ 0° AOI

S2:SINGLE LAYER MgF2

#### 4. FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT









Τ

LONGPASS FILTER			_			Edmund Ontice	C®
REV A	S1	S2					3
SHAPE	PLANO	PLANO				Ø25mm 450nm HIGH PERFORMANC	F
SURFACE QUALITY	40-20	40-20	THIRD ANGLE		TITLE	I ONGPASS FILTER	
CLEAR APERTURE	>80%	>80%		1			CUEFT
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN	mm	DWG NO	62982	1 OF 1

- 1. SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)
  - \$1: HARD DIELECTRIC SPUTTERED T(avg): ≥91% FROM 508 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 490nm @ 0° AOI T(abs): =50% FOR 500±5nm @ 0° AOI

S2:SINGLE LAYER MgF2

### 4. FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT







6



Τ

			_			Edmund (	)ntice®
REV A	S1	\$2					plics
SHAPE	PLANO	PLANO				Ø25mm 500nm HIGH PERI	
SURFACE QUALITY	40-20	40-20	THIRD ANGLE . PROJECTION	$\odot$	TITLE	I ONGPASS FILTE	R
CLEAR APERTURE	>80%	>80%		1			
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN	mm	DWG NO	62983	1 OF 1

- 1. SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)
  - \$1: HARD DIELECTRIC SPUTTERED T(avg): ≥91% FROM 560 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 539nm @ 0° AOI T(abs): =50% FOR 550±5.5nm @ 0° AOI

S2:SINGLE LAYER MgF2

#### 4. FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT









Τ

LONGPASS FILTER						Edmund Ontice®	
REV A	S1	S2					
SHAPE	PLANO	PLANO				Ø25mm 550nm HIGH PERFORMANCE	
SURFACE QUALITY	40-20	40-20	THIRD ANGLE		TITLE	LONGPASS FILTER	
CLEAR APERTURE	>80%	>80%		1			
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN	mm	DWG NO	62984 3HEET 1 OF 1	

- 1. SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)
  - \$1: HARD DIELECTRIC SPUTTERED T(avg): ≥91% FROM 610 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 588nm @ 0° AOI T(abs): =50% FOR 600±6nm @ 0° AOI

S2:SINGLE LAYER MgF2

### 4. FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT







6



LONGPASS FILTER

Τ

			_			Edmund Ontice	<b>C</b> ®
REV A	S1	\$2					3
SHAPE	PLANO	PLANO		1.		Ø25mm 600nm HIGH PERFORMANC	:F
SURFACE QUALITY	40-20	40-20	THIRD ANGLE . PROJECTION	THIRD ANGLE		LONGPASS FILTER	
CLEAR APERTURE	>80%	>80%		1			011557
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN	mm	DWG NO	62985	1 OF 1

- 1. SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)
  - \$1: HARD DIELECTRIC SPUTTERED T(avg): ≥91% FROM 660 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 637nm @ 0° AOI T(abs): =50% FOR 650±6.5nm @ 0° AOI

S2:SINGLE LAYER MgF2

#### 4. FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT







6



Τ

						Edmund Ontice	R
REV A	S1	\$2					
SHAPE	PLANO	PLANO				Ø25mm 650nm HIGH PERFORMANCE	
SURFACE QUALITY	40-20	40-20	THIRD ANGLE		TITLE	I ONGPASS FILTER	
CLEAR APERTURE	>80%	>80%		1			ICCT
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN	mm	DWG NO	62986 SH	DF 1

- 1. SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)
  - \$1: HARD DIELECTRIC SPUTTERED T(avg): ≥91% FROM 710 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 686nm @ 0° AOI T(abs): =50% FOR 700±7nm @ 0° AOI

S2:SINGLE LAYER MgF2

### 4. FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT







6



Τ

						Edmund Ontice	R
REV A	S1	\$2					2
SHAPE	PLANO	PLANO				Ø25mm 700nm HIGH PERFORMANCE	F
SURFACE QUALITY	40-20	40-20	THIRD ANGLE		TITLE	LONGPASS FILTER	
CLEAR APERTURE	>80%	>80%		1			CLIEFT
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN	mm	DWG NO	62987	1 OF 1

- 1. SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)
  - \$1: HARD DIELECTRIC SPUTTERED T(avg): ≥91% FROM 765 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 735nm @ 0° AOI T(abs): =50% FOR 750±7.5nm @ 0° AOI

S2:SINGLE LAYER MgF2

#### 4. FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT







6



					Edmund Ontice	<b>C</b> ®
REV A	S1	\$2				3
SHAPE	PLANO	PLANO			Ø25mm 750nm HIGH PERFORMANC	.F
SURFACE QUALITY	40-20	40-20		TITLE	LONGPASS FILTER	
CLEAR APERTURE	>80%	>80%				CLIEFT
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN mm	DWG NO	66234	1 OF 1

- 1. SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)
  - \$1: HARD DIELECTRIC SPUTTERED T(avg): ≥91% FROM 815 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 785nm @ 0° AOI T(abs): =50% FOR 800±8nm @ 0° AOI

S2:SINGLE LAYER MgF2

### 4. FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT







6



LONGPASS FILTER

Τ

			_			Edmund Ontice®
REV A	S1	S2				
SHAPE	PLANO	PLANO				Ø25mm 800nm HIGH PERFORMANCE
SURFACE QUALITY	40-20	40-20	THIRD ANGLE		TITLE	LONGPASS FILTER
CLEAR APERTURE	>80%	>80%				
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN	mm	DWG NO	66235 SHEET 1 OF 1

- 1. SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)
  - \$1: HARD DIELECTRIC SPUTTERED T(avg): ≥91% FROM 865 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 835nm @ 0° AOI T(abs): =50% FOR 850±8.5nm @ 0° AOI

S2:SINGLE LAYER MgF2

#### 4. FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT









Τ

LONGPASS FILTER						Edmund Ontic	с®
REV A	S1	S2					5
SHAPE	PLANO	PLANO		1		Ø25mm 850nm HIGH PERFORMAN	CF
SURFACE QUALITY	40-20	40-20	THIRD ANGLE -			I ONGPASS FILTER	
CLEAR APERTURE	>80%	>80%					CULLET
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN	mm	DWG NO	66236	1 OF 1

- 1. SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)
  - \$1: HARD DIELECTRIC SPUTTERED T(avg): ≥91% FROM 915 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 880nm @ 0° AOI T(abs): =50% FOR 900±9nm @ 0° AOI

S2:SINGLE LAYER MgF2

### 4. FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT







6



Τ

						Edmund Ontice®	)
REV A	S1	\$2					
SHAPE	PLANO	PLANO		1		Ø25mm 900nm HIGH PERFORMANCE	
SURFACE QUALITY	40-20	40-20	THIRD ANGLE _ PROJECTION	$\bigcirc \bigcirc$	TITLE	LONGPASS FILTER	
CLEAR APERTURE	>80%	>80%		1			
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN	mm	DWG NO	66237 SHE	.EI /F 1

- 1. SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)
  - \$1: HARD DIELECTRIC SPUTTERED T(avg): ≥91% FROM 965 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 930nm @ 0° AOI T(abs): =50% FOR 950±9.5nm @ 0° AOI

S2:SINGLE LAYER MgF2

### 4. FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT







6



Τ

						Edmund Ontice	R
REV A	S1	S2					
SHAPE	PLANO	PLANO		I		Ø25mm, 950nm, HIGH PEREORMANCE	:
SURFACE QUALITY	40-20	40-20	THIRD ANGLE _ PROJECTION			LONGPASS FILTER	
CLEAR APERTURE	>80%	>80%		1			
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN	mm	DWG NO	66238	I OF 1

- 1. SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)
  - \$1: HARD DIELECTRIC SPUTTERED
    T(avg): ≥91% FROM 1020 1650nm @ 0° AOI
    T(avg): ≤0.01% FROM 200 980nm @ 0° AOI
    T(abs): =50% FOR 1000±10nm @ 0° AOI

S2:SINGLE LAYER MgF2

#### 4. FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT







6



LONGPASS FILTER

Τ

			_			Edmund Ontice®
REV A	S1	S2				
SHAPE	PLANO	PLANO				Ø25mm 1000nm HIGH PEREORMANCE
SURFACE QUALITY	40-20	40-20	THIRD ANGLE PROJECTION	$\bigcirc \bigcirc \bigcirc$	TITLE	I ONGPASS FILTER
CLEAR APERTURE	>80%	>80%		1		
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN	mm	DWG NO	66239 SHEEL 1 OF 1

- 1. SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)
  - S1: HARD DIELECTRIC SPUTTERED T(avg): ≥91% FROM 1070 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 1030nm @ 0° AOI T(abs): =50% FOR 1050±10.5nm @ 0° AOI

S2:SINGLE LAYER MgF2

#### 4. FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT









Τ

LONGPASS FILTER						Edmund Ontic	<b>C</b> ®	
REV A	S1	S2					5	
SHAPE	PLANO	PLANO				Ø25mm 1050nm HIGH PERFORMANCE		
SURFACE QUALITY	40-20	40-20			TITLE	LONGPASS FILTER		
CLEAR APERTURE	>80%	>80%		1			CLIEFT	
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN	mm	DWG NO	66240	1 OF 1	

- 1. SUBSTRATE
- 2. SURFACE S2 TO BE PARALLEL WITH SURFACE S1 TO WITHIN 5 arcsec
- 3. COATING (APPLY ACROSS COATING APERTURE)
  - S1: HARD DIELECTRIC SPUTTERED T(avg): ≥91% FROM 1120 - 1650nm @ 0° AOI T(avg): ≤0.01% FROM 200 - 1080nm @ 0° AOI T(abs): =50% FOR 1100±11nm @ 0° AOI

S2:SINGLE LAYER MgF2

#### 4. FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. TRANSMITTED WAVEFRONT DISTORTION, RMS: ≤λ/4 @ 633nm
- 7. ROHS COMPLIANT







6



LONGPASS FILTER

Τ

					Edmund Ontice®
REV A	S1	S2			
SHAPE	PLANO	PLANO			Ø25mm 1100nm HIGH PEREORMANCE
SURFACE QUALITY	40-20	40-20		TITLE	LONGPASS FILTER
CLEAR APERTURE	>80%	>80%			
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN mm	DWG NO	66241 SHEEI 1 OF 1