

NOTES:

1. SUBSTRATE: VACUUM UV GRADE CaF2

- 2. SURFACE S2 TO BE PARALLEL WITH \$1 TO <1 arcmin
- 3. COATING (APPLY ACROSS CLEAR APERTURE) \$1 & \$2: R(ABS) ≤ 0.25% AT 880nm @ 0° AOI R(AVG) ≤ 1.25% FROM 400-870nm @ 0° AOI R(AVG) ≤ 1.25% FROM 890-1000nm @ 0° AOI







FOR INFORMATION ONLY: DO NOT MANUFACTURE PARTS TO THIS DRAWING

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

	\$1	\$2		
SHAPE	PLANO	PLANO		
SURFACE QUALITY	40-20	40-20	THI PR	
CLEAR APERTURE	Ø45.00	Ø45.00	_	
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	Α	

		15	
THIRD ANGLE - PROJECTION		TITLE	Q
ALL DIMS IN	mm	DWG NO	148

Ø50mm, VIS - NIR, CALCIUM FLUORIDE WINDOW

Edmund Optics®

mm DWG NO 14821 SHEET 1 OF 1